



MATHESON

ask...The Gas Professionals™

MATHESON

166 Keystone Drive
 Montgomeryville, PA 18936
 +1 (800) 828-4313
www.mathesongas.com

UNITS	: SLPM CHLORINE	DATE	: 03/06/95 (F)
TUBE NUMBER	: #6 (E606)	STD CONDITIONS	: 1 ATMOS. & 70 DEG F
SERIAL NUMBER	: TYPICAL	GAS TEMPERATURE	: 75 DEG F
FLOAT MATERIAL:	316 ST. STL.	PRESSURE IN TUBE	: 100 PSIG

SCALE READING	FLOW RATE	SCALE READING	FLOW RATE	SCALE READING	FLOW RATE	SCALE READING	FLOW RATE
150.0	12.31	110.0	9.44	70.0	6.35	30.0	3.01
149.0	12.24	109.0	9.36	69.0	6.27	29.0	2.93
148.0	12.17	108.0	9.29	68.0	6.19	28.0	2.84
147.0	12.10	107.0	9.22	67.0	6.11	27.0	2.76
146.0	12.03	106.0	9.14	66.0	6.03	26.0	2.67
145.0	11.96	105.0	9.07	65.0	5.94	25.0	2.59
144.0	11.88	104.0	8.99	64.0	5.86	24.0	2.50
143.0	11.81	103.0	8.92	63.0	5.78	23.0	2.42
142.0	11.74	102.0	8.84	62.0	5.70	22.0	2.33
141.0	11.67	101.0	8.77	61.0	5.62	21.0	2.25
140.0	11.60	100.0	8.69	60.0	5.54	20.0	2.16
139.0	11.53	99.0	8.62	59.0	5.45	19.0	2.08
138.0	11.46	98.0	8.54	58.0	5.37	18.0	1.99
137.0	11.39	97.0	8.47	57.0	5.29	17.0	1.91
136.0	11.32	96.0	8.39	56.0	5.20	16.0	1.82
135.0	11.25	95.0	8.31	55.0	5.12	15.0	1.74
134.0	11.17	94.0	8.24	54.0	5.04	14.0	1.65
133.0	11.10	93.0	8.16	53.0	4.96	13.0	1.57
132.0	11.03	92.0	8.08	52.0	4.87	12.0	1.48
131.0	10.96	91.0	8.01	51.0	4.79	11.0	1.40
130.0	10.89	90.0	7.93	50.0	4.70	10.0	1.31
129.0	10.82	89.0	7.85	49.0	4.62	9.0	1.23
128.0	10.75	88.0	7.78	48.0	4.54	8.0	1.15
127.0	10.67	87.0	7.70	47.0	4.45	7.0	1.06
126.0	10.60	86.0	7.62	46.0	4.37	6.0	0.98
125.0	10.53	85.0	7.54	45.0	4.29	5.0	0.89
124.0	10.46	84.0	7.46	44.0	4.20		
123.0	10.39	83.0	7.39	43.0	4.12		
122.0	10.31	82.0	7.31	42.0	4.03		
121.0	10.24	81.0	7.23	41.0	3.95		
120.0	10.17	80.0	7.15	40.0	3.86		
119.0	10.10	79.0	7.07	39.0	3.78		
118.0	10.02	78.0	6.99	38.0	3.69		
117.0	9.95	77.0	6.91	37.0	3.61		
116.0	9.88	76.0	6.83	36.0	3.52		
115.0	9.81	75.0	6.75	35.0	3.44		
114.0	9.73	74.0	6.67	34.0	3.35		
113.0	9.66	73.0	6.59	33.0	3.27		
112.0	9.59	72.0	6.51	32.0	3.18		
111.0	9.51	71.0	6.43	31.0	3.10		