



MATHESON

ask...The Gas Professionals™

MATHESON

166 Keystone Drive
 Montgomeryville, PA 18936
 +1 (800) 828-4313
www.mathesongas.com

UNITS : SLPM NITROUS OXIDE DATE : 08-01-71
 TUBE NUMBER : 604 (E700) STD CONDITIONS : 1 ATMOS. & 70 DEG F
 SERIAL NUMBER : TYPICAL GAS TEMPERATURE : 70 DEG F
 FLOAT MATERIAL: 316 ST. STL. PRESSURE IN TUBE : 0 PSIG

E703NOXI.DAT

SCALE READING	FLOW RATE	SCALE READING	FLOW RATE	SCALE READING	FLOW RATE	SCALE READING	FLOW RATE
150.0	14.59	110.0	10.56	70.0	6.64	30.0	2.72
149.0	14.48	109.0	10.46	69.0	6.54	29.0	2.62
148.0	14.38	108.0	10.36	68.0	6.44	28.0	2.53
147.0	14.28	107.0	10.26	67.0	6.34	27.0	2.43
146.0	14.17	106.0	10.16	66.0	6.25	26.0	2.33
145.0	14.07	105.0	10.06	65.0	6.15	25.0	2.24
144.0	13.97	104.0	9.97	64.0	6.05	24.0	2.14
143.0	13.86	103.0	9.87	63.0	5.95	23.0	2.04
142.0	13.76	102.0	9.77	62.0	5.85	22.0	1.94
141.0	13.66	101.0	9.67	61.0	5.76	21.0	1.85
140.0	13.55	100.0	9.57	60.0	5.66	20.0	1.75
139.0	13.45	99.0	9.48	59.0	5.56	19.0	1.65
138.0	13.35	98.0	9.38	58.0	5.46	18.0	1.56
137.0	13.25	97.0	9.28	57.0	5.36	17.0	1.46
136.0	13.15	96.0	9.18	56.0	5.27	16.0	1.37
135.0	13.05	95.0	9.08	55.0	5.17	15.0	1.27
134.0	12.95	94.0	8.99	54.0	5.07	14.0	1.17
133.0	12.84	93.0	8.89	53.0	4.97	13.0	1.08
132.0	12.74	92.0	8.79	52.0	4.87	12.0	0.98
131.0	12.64	91.0	8.69	51.0	4.78	11.0	0.89
130.0	12.54	90.0	8.59	50.0	4.68	10.0	0.79
129.0	12.44	89.0	8.50	49.0	4.58	9.0	0.70
128.0	12.34	88.0	8.40	48.0	4.48	8.0	0.60
127.0	12.24	87.0	8.30	47.0	4.38	7.0	0.50
126.0	12.14	86.0	8.20	46.0	4.29	6.0	0.41
125.0	12.04	85.0	8.10	45.0	4.19	5.0	0.32
124.0	11.94	84.0	8.01	44.0	4.09		
123.0	11.84	83.0	7.91	43.0	3.99		
122.0	11.74	82.0	7.81	42.0	3.89		
121.0	11.64	81.0	7.71	41.0	3.80		
120.0	11.54	80.0	7.62	40.0	3.70		
119.0	11.44	79.0	7.52	39.0	3.60		
118.0	11.35	78.0	7.42	38.0	3.50		
117.0	11.25	77.0	7.32	37.0	3.40		
116.0	11.15	76.0	7.22	36.0	3.31		
115.0	11.05	75.0	7.13	35.0	3.21		
114.0	10.95	74.0	7.03	34.0	3.11		
113.0	10.85	73.0	6.93	33.0	3.01		
112.0	10.75	72.0	6.83	32.0	2.92		
111.0	10.65	71.0	6.73	31.0	2.82		