



Matheson Tri-Gas, Inc.
909 Lake Carolyn Parkway
Irving, TX 75039

Medical Grade Oxygen

This certifies that all product labeled as Oxygen, USP provided by Matheson Tri-Gas meets and exceeds the USP monograph for Medical Grade Oxygen.

CHARACTERISTIC	CGA G-4.3 QVL A	UNITED STATES PHARMACOPEIA (USP)	MATHESON MEDICAL GRADE
	MEDICAL		
Oxygen (min)	99.0 %	NLT 99.0 %	≥ 99.2 %
Carbon Dioxide	300 ppm ^a	NMT 0.03 ppm ^a	n/a
Carbon Monoxide	10 ppm ^a	NMT 0.001 ppm ^a	n/a
Odor	None	None	None
Identity	Not specified	Pass	Pass

^a Oxygen produced by the air-liquefaction process is exempt from the requirements of the tests for Carbon Dioxide and Carbon Monoxide.

Certified By (if applicable):

Date: